

26RFP002 Mobile Health Vehicle

Addendum #1 Questions and Answers

Issued: 01/23/2026

1 of 3

Question Number	Date Received	Question	Answer
Q1-1	01/13/2026 1:51 PM	Can you say if you're more looking for a VAN sized vehicle (like the Ram Promaster 2500), or are you seeking a much larger vehicle built on a Cab/Chassis with a roomier housing?	<i>SNHD prefers a van like the RAM Promaster 2500 or equivalent.</i>
Q1-2	01/13/2026 1:51 PM	Do you have a sketch/draft layout of where internal furnishings might be placed to work best for SNHD?	<i>SNHD does not have a preferred floorplan.</i>
Q1-3	01/13/2026 1:51 PM	For ADA compliance, are you thinking a wheelchair entry aid is what you need?	<i>RFP Section II.B.13 has been revised to: Include optional rear mounted wheelchair lift. Attachment B has also been revised to include this option and is included here.</i>
Q1-4	01/13/2026 1:51 PM	Do you have the Model Number of the Full Size Exam Bed you have in mind?	<i>SNHD does not have a preferred make/model number for the full size exam bed. Please provide a built list of the proposed equipment.</i>
Q2-1	09/19/2025 3:47 PM	Is it possible for SNHD to provide the vehicle shell to the contractor to complete the conversion?	<i>SNHD seeks a complete, fully equipped mobile health vehicle. SNHD will not provide the vehicle shell to the Contractor to complete the conversion.</i>
Q3-1	1/16/2026 1:57 PM	Do you have a floor plan/layout diagram in mind for this project?	<i>SNHD does not have a preferred floorplan.</i>
Q3-2	1/16/2026 1:57 PM	What services will you be offering? Mainly primary care? Any behavioral health?	<i>Primary care, vaccines and STI screening, testing, and treatment to vulnerable and underserved populations.</i>
Q3-3	1/16/2026 1:57 PM	Do you need a private room with accordion door?	<i>No.</i>
Q3-4	1/16/2026 1:57 PM	Do you need exam light for treatment room?	<i>No.</i>
Q3-5	1/16/2026 1:57 PM	Do you need a handicap lift inside treatment room? Braun /Ricon	<i>No.</i>
Q3-6	1/16/2026 1:57 PM	How many staff working inside the mobile clinic?	<i>Two to three.</i>
Q3-7	1/16/2026 1:57 PM	Do you need a bathroom with black tank and sink inside?	<i>No.</i>
Q3-8	1/16/2026 1:57 PM	Do you need a blood draw chair?	<i>No.</i>

26RFP002 Mobile Health Vehicle

Addendum #1 Questions and Answers

Issued: 01/23/2026

2 of 3

Q3-9	1/16/2026 1:57 PM	Do you need another sink for treatment room?	No.
Q3-10	1/16/2026 1:57 PM	How many total sinks inside mobile do you prefer?	<i>One sink.</i>
Q3-11	1/16/2026 1:57 PM	How many upper cabinets do you need?	<i>The number of upper cabinets should be what best fits the proposed space/layout.</i>
Q3-12	1/16/2026 1:57 PM	Would you like a sliding window near desk area?	No.
Q3-13	1/16/2026 1:57 PM	Do you need a mobile lab area with sound curtain?	No.
Q3-14	1/16/2026 1:57 PM	Do you need a locking pass thru door from box to chassis?	<i>See RFP section II.B.16.b.</i>
Q3-15	1/16/2026 1:57 PM	Will you consider a 26 Ft new clinic built on a Ford E450 chassis?	<i>SNHD will consider alternatives to RAM Promaster 2500 provided the proposed vehicle meets all other RFP requirements.</i>
Q3-16	1/16/2026 1:57 PM	Do you need electric wall heaters when using during cold winter months?	<i>No. RFP Section II.B.5 has been revised to: Include rooftop HVAC system to provide internal heating and cooling.</i>
Q3-17	1/16/2026 1:57 PM	Will you consider two new mobile unit options?	<i>Per the RFP, "Contractor may propose up to three (3) different MHV designs."</i>
Q4-1a	1/19/2026 10:16 AM	RAM 2500 Promaster or equivalent: Will the vehicle require 4-wheel drive for off the road services?	No.
Q4-1b	1/19/2026 10:16 AM	Besides a van platform, will you also accept a wider body platform providing adequate space for the items you have specified and room for patients and operators?	<i>SNHD will consider wider body platforms provided the proposed vehicle meets all other RFP requirements.</i>
Q4-2	1/19/2026 10:16 AM	Electrical: You specify the generator and shore power as options. What provisions do you plan to have to operate the 120-volt outlets, roof top air conditioner, vaccine refrigerator, heaters and other 120-volt items if those options are not selected for the build?	<i>RFP Section II.B.8 has been revised to: Include optional 5kW generator, if available. Attachment B has also been revised to include this option and is attached.</i> <i>Also, RFP Section II.B includes the following new requirement: Include a 2kW Inverter with a second battery.</i>
Q4-3	1/19/2026 10:16 AM	ADA Compliant: Are you looking for a ramp or an ADA compliant wheelchair lift at the rear or side door? Will the unit be transporting a person in a wheelchair requiring ADA wheelchair tie downs?	<i>RFP Section II.B.13 has been revised to: Include optional rear mounted wheelchair lift. Attachment B has also been revised to include this option and is included here.</i>

26RFP002 Mobile Health Vehicle**Addendum #1 Questions and Answers****Issued: 01/23/2026****3 of 3**

Q4-4	1/19/2026 10:16 AM	Graphics wrap: Do you require a full vehicle wrap or a partial wrap?	<i>The wrap will be outdoor UV-stabilized material, approximately 200 sq. ft., applied to both sides, front and back of vehicle, in a single color (SNHD blue).</i>
Q4-5	1/19/2026 10:16 AM	MHV must include: Do you plan to include a phlebotomy chair or position for blood draws?	<i>No.</i>
Q5-1	1/20/2026 9:53 AM	Will SNHD consider proposals on a different/larger platform?	<i>SNHD will consider different/larger platforms provided the proposed vehicle meets all other RFP requirements.</i>
Q5-2	1/20/2026 9:53 AM	What dimensions do you consider a full-size desk?	<i>Nominally 30" - 36" W, 16" - 20" D, and standard desk height depending on layout.</i>
Q5-3	1/20/2026 9:53 AM	What size of vaccine refrigerator does this unit need?	<i>Undercounter type to fit the space, nominally 23" H x 17" W x 16" D. Dometic brand or equivalent.</i>
Q5-4	1/20/2026 9:53 AM	Do you have a preferred fresh and wastewater tank capacity, or should we propose our recommended?	<i>Propose Contractor recommended.</i>
Q6-1	09/23/2025 3:25 PM	Will SNHD consider as an alternative to the requested RAM ProMaster [e.g., Ford E-450 chassis]?	<i>SNHD will consider alternatives to RAM Promaster 2500 provided the proposed vehicle meets all other RFP requirements.</i>

Revised Attachment B **Price Form**

The prices to complete this work shall include all labor, materials, tools, supplies, equipment and services necessary to comply with the specifications, terms and conditions set forth in this RFP. No additional costs may be billed without prior SNHD approval. The prices shall be valid for 90 days.

Description of Mobile Health Vehicle or Demo Vehicle being provided:

Model Year: _____

Odometer Mileage: _____ (Demo vehicle only)

Brand: _____

Model: _____

Length: _____

Weight: _____

Delivery Date: _____

Other Information:

Price for NEW Mobile Health Vehicle: _____

Price for DEMO Vehicle (indicate discount): _____

Optional electric awning: _____

Optional 5kW generator: _____

Optional 30-amp shore power service/receptacles/cable: _____

Optional rear mounted wheelchair lift: _____

Delivery to 280 S Decatur Blvd., Las Vegas, NV 89107: _____

Company Name: _____

Authorized Signature: _____ Date: _____

Printed Name and Title: _____

Phone: _____ Email: _____